

Certificate of Compliance



No. OP170601.QASUS08

Technical Construction File no. TLEJS201705271876-2

Certificate's
Holder:

Qingdao Antech Scientific Co., Ltd.
No. 11 Zaoyuan Road, Qingdao, Shandong, China

Certification
ECM Mark:



Product:
Model(s):

Medical / Laboratory Refrigerator
(see the following annex I)

Verification to:

Standard:
EN 61000-3-2:2014, EN 61000-3-3:2013,
EN 55014-1:2006+A2:2011,
EN 55014-2:1997+A1:2001,
EN 60335-2-24:2010, EN 60335-1:2012+A11:2014

related to CE Directive(s):
2014/35/EU (Low Voltage)
2014/30/EU (Electromagnetic Compatibility)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** and to perform all the necessary activities, as required by the Directive before placing the **CE** Mark on the product(s).

Date of issue 01 June 2017

Expiry date 31 May 2022

Chief Manager
Marta Morina

Deputy Manager
Amanda Payne

Annex I



No. OP170601.QASUS08

Technical Construction File no. TLEJS201705271876-2

Model(s):

MBR-108/158/208/258/258S/358/358S/458/658/858/1008/1308/1508
MPR-60/62/68/110/116/140/150/260/296/306/360/368/406/426/626/640/760
/1006/1206/1506
MPR-60S/62S/68S/110S/116S/140S/260S/296S/306S/360S/368S/406S/426S/626S
/640S/760S/1006S/1206S/1506S
MBR-30M/80M
ATCB-70/130/200/120/200
MBA-5, MBA-10
BC-120/240
ATCB-70/130/200
ATCF-120/200
MCR-10, MCR-20, MCR-30, MCR-40, MCR-50, MCR-60, MCR-80, MCR-90,
MCR-100
MVR-120/240/300/400/500
SVR-60/100/200/300/400