



Product Service

# Attestation of Conformity


No. N8A 17 08 89832 005

**Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD**No.2 Weixin Road, SIP  
215122 Suzhou, Jiangsu  
PEOPLE'S REPUBLIC OF CHINA**Product: Laboratory Equipment  
Bio. Safety Cabinet****Model(s): BSC-1003IIA2, BSC-1003IIA2-S, BSC-1003IIA2-E  
BSC-1003IIA2-SE, BSC-1303IIA2, BSC-1303IIA2-S  
BSC-1303IIA2-E, BSC-1303IIA2-SE, BSC-1603IIA2  
BSC-1603IIA2-S, BSC-1603IIA2-E, BSC-1603IIA2-SE**

<b>Parameters:</b>	Rated input voltage:	220-240VAC
	Rated input frequency:	50/60Hz
	Max. power:	1.5kVA(BSC-1003IIA2 BSC-1003IIA2-S, BSC-1003IIA2-E BSC-1003IIA2-SE, BSC-1303IIA2 BSC-1303IIA2-S, BSC-1303IIA2-E BSC-1303IIA2-SE) 1.8kVA(BSC-1603IIA2 BSC-1603IIA2-S, BSC-1603IIA2-E BSC-1603IIA2-SE)
	Class of MSC:	II
	Electrical protection class:	I

**Tested according to:** EN 12469:2000  
EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

**Test report no.:** 884001506302**Date,** 2017-09-06
  
( Bin Li )


**CE** After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1