



Attestation of Conformity

No. E8A 17 07 89832 001

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Name of Object: Laboratory Equipment
Bio. Safety Cabinet

Model(s): BSC-1003IIA2, BSC-1303IIA2, BSC-1603IIA2
BSC-1003IIA2-S, BSC-1303IIA2-S, BSC-1603IIA2-S
BSC-1003IIA2-E, BSC-1303IIA2-E, BSC-1603IIA2-E
BSC-1003IIA2-SE, BSC-1303IIA2-SE, BSC-1603IIA2-SE

Description of Object:

Rated voltage: 220-240VAC, 50/60Hz
Max. power consumption: 1.5kW (BSC-1003IIA2, BSC-1303IIA2
BSC-1003IIA2-S, BSC-1303IIA2-S
BSC-1003IIA2-E, BSC-1303IIA2-E
BSC-1003IIA2-SE, BSC-1303IIA2-SE)
1.8kW (BSC-1603IIA2, BSC-1603IIA2-S
BSC-1603IIA2-E, BSC-1603IIA2-SE)
Rated input voltage: 220-240VAC, 50/60Hz
Class of MSC: II
Protection class: I

Tested according to: EN 61326-1:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 888001506402

Date, 2017-08-10

(Jing Zhang)



After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. That declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.