



Product Service

Attestation of Conformity

No. N8A 089832 0011 Rev. 00

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Product: Laboratory Equipment
Clean Bench

Model(s): HS-840, HS-840-U, HS-1300, HS-1300-U,
HS-1800H-U, VS-840K, VS-840K-U,
VS-1300L, VS-1300L-U, VS-1800L-U,
HHS-1000S-U, HHS-1300S-U, HHS-1600S-U,
HHS-1800S-U, HVS-1000S-U, HVS-1300S-U,
HVS-1600S-U, HVS-1800S-U


| | | |
|--------------------|-----------------------|---|
| Parameters: | Rated voltage: | 220-240V~ |
| | Rated frequency: | 50/60Hz |
| | Rated power: | 1300VA (HS-840, HS-840-U, VS-840K, VS-840K-U, HHS-1000S-U, HHS-1300S-U, HHS-1600S-U, HHS-1800S-U, HVS-1000S-U, HVS-1300S-U, HVS-1600S-U, HVS-1800S-U); 1600VA (HS-1800H-U, VS-1800L-U, HS-1300, HS-1300-U, VS-1300L, VS-1300L-U) |
| | Protection class: | I |
| | Degree of protection: | IP20 |

Tested according to: EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 4842018105500

Date, 2018-12-03


(Bin Li)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證書 ◆ CERTIFICADO ◆ CERTIFICAT

A4 / 07.17