

# Certificate of Compliance



No. OP170601.QASUS08

Technical Construction File no. TLEJS201705271876-2

Certificate's  
Holder:

Qingdao Antech Scientific Co., Ltd.  
No. 11 Zaoyuan Road, Qingdao, Shandong, China

Certification  
ECM Mark:



Product:  
Model(s):

Medical / Laboratory Refrigerator  
(see the following annex I)

Verification to:

Standard:  
EN 61000-3-2:2014, EN 61000-3-3:2013,  
EN 55014-1:2006+A2:2011,  
EN 55014-2:1997+A1:2001,  
EN 60335-2-24:2010, EN 60335-1:2012+A11:2014

related to CE Directive(s):  
2014/35/EU (Low Voltage)  
2014/30/EU (Electromagnetic Compatibility)

**Remark:** The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at [www.entecerma.it](http://www.entecerma.it). This Certificate of Compliance can be checked for validity at [www.entecerma.it](http://www.entecerma.it)

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** and to perform all the necessary activities, as required by the Directive before placing the **CE** Mark on the product(s).

Date of issue 01 June 2017

Expiry date 31 May 2022

Chief Manager  
Marta Morina

Deputy Manager  
Amanda Payne

# Annex I



No. OP170601.QASUS08

Technical Construction File no. TLEJS201705271876-2

## Model(s):

MBR-108/158/208/258/258S/358/358S/458/658/858/1008/1308/1508  
MPR-60/62/68/110/116/140/150/260/296/306/360/368/406/426/626/640/760  
/1006/1206/1506  
MPR-60S/62S/68S/110S/116S/140S/260S/296S/306S/360S/368S/406S/426S/626S  
/640S/760S/1006S/1206S/1506S  
MBR-30M/80M  
ATCB-70/130/200/120/200  
MBA-5, MBA-10  
BC-120/240  
ATCB-70/130/200  
ATCF-120/200  
MCR-10, MCR-20, MCR-30, MCR-40, MCR-50, MCR-60, MCR-80, MCR-90,  
MCR-100  
MVR-120/240/300/400/500  
SVR-60/100/200/300/400